DCB0129 Clinical Safety Case Report
Suvera Recall

The Clinical Safety Case Report (CSCR) summarises the clinical safety management activities, evidence and residual risks for the Suvera Recall module, supporting assessment and authorisation before the product goes live.

Document Management
Version: 1.0 — Draft
Date: 14 January 2026
Document Owner: Clinical Safety Officer (Head of Product Safety, CSO)
Classification: Internal

Confidentiality Statement
This document contains information that is privileged, confidential, or otherwise exempt from disclosure. It is intended for the use of the individual or entity to which it is addressed. If you are not the intended recipient, you are hereby notified that any dissemination, distribution or copying of this document is strictly prohibited.

Document Management
Revision History

Version	Date	Summary of Changes	Author
0.1	Dec 15, 2025	Initial draft created from DCB0129 template; Changes 8 workflows in multi-step	Priyanka Bhimani
0.2	Dec 19, 2025	Updated Appointment and Recall Workflow; Added the Clinical Safety Case Report Template; Updated Risk Management section	Priyanka Bhimani
0.3	Dec 22, 2025	Updated hazard assessment tables; Added Risk Matrix; Updated Risk Register; Updated Risk Management Plan	Priyanka Bhimani
0.4	Dec 29, 2025	Updated hazard assessment tables; Updated Risk Register; Updated Risk Management Plan	Priyanka Bhimani
0.5	Jan 5, 2026	Updated hazard assessment tables; Updated Risk Register; Updated Risk Management Plan	Priyanka Bhimani
0.6	Jan 8, 2026	Updated hazard assessment tables; Updated Risk Register; Updated Risk Management Plan	Priyanka Bhimani
0.7	Jan 10, 2026	Updated hazard assessment tables; Updated Risk Register; Updated Risk Management Plan	Priyanka Bhimani
0.8	Jan 12, 2026	Updated hazard assessment tables; Updated Risk Register; Updated Risk Management Plan	Priyanka Bhimani
0.9	Jan 13, 2026	Updated hazard assessment tables; Updated Risk Register; Updated Risk Management Plan	Priyanka Bhimani
1.0	Jan 14, 2026	Initial release for approval	Priyanka Bhimani
1.1	Jan 20, 2026	Updated post-deployment monitoring plan	Priyanka Bhimani
1.2	Jan 22, 2026	Updated risk assessment	Ryan Flagg

Contents

Confidentiality Notice
Document Management
Introduction
System Definition / Overview
Aims (Intended Use)
Clinical Context
Users / Scale
System Functionality Overview
Software Identification
Software Boundaries
Deployment and Technical Architecture
Scope Justification
Clinical Risk Management System
Clinical Risk Management Plan
Clinical Risk Assessment
Clinical Risk Register
Clinical Risk Management Activities
Post-Deployment Monitoring
Safety Incidents
Conclusion
Appendix A
Definitions
Risk Matrix
Risk Acceptability Criteria

Abbreviations & Definitions

Term	Definition
CSCR	Clinical Safety Case Report
CSO	Clinical Safety Officer
AHCP	Accredited Health Record
DCB0129	Clinical Risk Management: its Application in the Manufacture of Health IT Systems
DPI	Long Term Conditions
PPI	Patient Portal
PRO	Patient Records

Introduction
The purpose of this Clinical Safety Case Report is to provide the structured clinical safety evidence and rationale required to support the safe deployment of the Suvera Recall module. It aims to accurately define the system's boundaries, functionality, and intended use.

It demonstrates that:

- all risks associated with the system have been identified
- appropriate risk controls have been put in place
- residual risks have been reduced to an acceptable level

System Definition / Overview
Suvera Recall is a cloud-based clinical recall module developed by Suvera to support NP practice teams in managing patient recalls, appointments, and follow-up pathways. It is designed to integrate with existing practice management systems and is intended for use by clinical and administrative staff in UK primary care settings.

The system is a web-based application and does not physically control devices. It is delivered as a software-as-a-service (SaaS) product and is accessed via standard web browsers on desktop and mobile devices.

Recall enables practice teams to view patient cohorts individually, proactively or reactively identify patients requiring follow-up, and manage appointments and communications. The system does not make clinical decisions autonomously; all clinical judgement remains with the user. Its role is limited to supporting workflow, reminders, and record-keeping.

Interoperability activity is documented but out-of-scope for the patient's electronic health record as the system is not integrated with any EHR.

The system is intended to support routine, non-urgent clinical follow-up and is not designed for emergency or time-critical use. Users are instructed that urgent clinical concerns must be escalated through established clinical pathways rather than relying on the system.

Suvera Recall is a standalone software system and does not control, configure, or interface with medical devices. It is delivered as a software-as-a-service (SaaS) product and is accessed via standard web browsers on desktop and mobile devices.

The system is a web-based application and does not physically control devices. It is delivered as a software-as-a-service (SaaS) product and is accessed via standard web browsers on desktop and mobile devices.

In Scope
The following functions are included within Suvera Recall and will be assessed:

- Creation and management of recall lists
- Automated and manual patient recall scheduling
- Appointment booking and rescheduling
- Clinician review and sign-off of recall outcomes
- Patient communications (SMS, email, letter)
- Recall status tracking and audit trail

Out of Scope
The following are explicitly excluded from this Clinical Safety Case Report:

- Clinical governance of practice clinicians making the recalls and primary care of patients
- Clinical decision-making regarding whether a patient should be recalled or not
- Integration with practice management systems (PMS) or electronic health records (EHR)
- Patient-facing appointment and online portal, clinician-facing view is direct
- Telephony systems used by practice staff to contact patients
- System-level views of third-party digital health providers used by practice staff
- Implementation of recalls
- Data warehousing or reporting
- Video conferencing or telephony platforms
- Prescribing functionality

Aims (Intended Use)
Clinical Purpose
Identification of patients with long-term conditions due to lapsed and patients requiring routine monitoring or follow-up appointments, and facilitation of timely recall.

Clinical Context
Recall is a tool for supporting practice teams in delivering the recall of patients who require routine monitoring or follow-up appointments. It is designed to support, not replace, clinical judgement.

Records will appear within their recall dashboard in the form of a live clinical list or task queue.

Recall is not intended for use in emergency or urgent clinical situations. Practice staff are advised that they should not rely on the system for time-critical clinical decisions.

Users / Scale
- Staff (GP, Clinical Pharmacists, Nurse, Administrators) at about 10 user practices
- NP practices

System Functionality Overview
Software Identification

Software Description
Suvera Recall is a web-based system, accessed via secure login, and is offered by Suvera through a SaaS subscription model.

Deployment and Technical Architecture

Features:

- Clients access the web application via modern browsers
- Backend services handle business logic, data processing, and integrations
- Database stores:
  - Patient demographics
  - Patients under Suvera care
  - Recall schedules and statuses
  - Appointment records
  - Audit logs
- Authentication service manages user access and permissions
- Email service sends automated and manual recall communications
- SMS service sends automated and manual recall communications
- Integrations:
  - Practice management system for appointment booking status
  - Patient records for patient demographics
  - Clinician review for clinical sign-off
  - Patient communications for recall messages
  - Audit trail for recall activity

Scope Justification Statement
Recall users interact with GPs, Clinical Pharmacists, Nurses, Administrators at each practice.

System dependencies
Integrations

- Email
  - Email is used - changes to patient medical records are sent from the EHR to Suvera Recall
  - Email is sent from Suvera to patients, but only if patient is in the EHR
- SMS
  - Text is sent from Suvera to patients, but only if patient is in the EHR
- Clinician review
  - Clinicians review patient's EHR records to support operational workflow
- Scheduling
  - Patients can book appointments via Suvera Recall
  - A call for multiple patients, either by bulk search or "one-off" call to a filtered list of patients

Third-party dependencies
Suvera Recall is dependent on third-party services for its SMS and EHR. The primary third-party dependencies are listed below. Failure or degradation of these services may affect the availability of Suvera Recall.

- Twilio - appointment scheduling system, not integrated with the EHR
- Google Workspace - appointment scheduling system, not integrated with the EHR
- Microsoft Azure - appointment scheduling system, not integrated with the EHR
- AWS (us-1) Control region
- Twilio - SMS messaging for patient recalls
- Twilio - SMS messaging for patient recalls
- Google Workspace - email messaging for patient recalls
- Integrated with the EHR

Clinical Risk Management System
Suvera Recall is developed and maintained under a formal clinical risk management system to ensure that risks to patient safety are identified, assessed, controlled, and reviewed throughout the product lifecycle.

This approach aims to ensure compliance with the clinical risk management standards applicable to digital health technologies, including DCB0129 and ISO 14971.

The Suvera team and its clinical safety officer are responsible for the creation and maintenance of the clinical risk management system, including the clinical risk management plan, clinical risk register, and post-deployment monitoring activities.

Team, Roles and Responsibilities

Role	Responsibilities
Clinical Safety Officer (CSO)	• Owns the Clinical Risk Management Plan and Clinical Risk Register • Leads and approves hazard identification and risk assessments • Determines whether residual risks are acceptable • Signs off the Clinical Safety Case Report • Oversees post-deployment monitoring and safety incidents
Ryan Flagg	• Engineering management and oversight • Technical risk controls
Ivan Nunley	• Software Release Manager

Clinical Risk Analysis
Clinical risks associated with the use of Suvera Recall have been made limited to the way the system is designed and the processes that support its use.

- Identification and analysis is results the hazards and primarily use the SWIFT methodology.
- Assessments of all hazards are documented in the Clinical Risk Register, including:
  - Severity
  - Likelihood
  - Initial risk score
  - Risk controls
  - Residual risk score
- Risks are reviewed and scored as high, medium or low.
- Risks are reduced to the lowest practicable level.
- Residual risks are reviewed by the Clinical Safety Officer (CSO).
- Only residual risks deemed acceptable by the CSO are considered suitable for release.

Clinical Risk Evaluation
Risk evaluation is carried out by the Clinical Safety Officer (CSO) and is based on the severity of potential harm to patients and the likelihood of that harm occurring.

Risks are assessed using a 5 × 5 risk matrix, with severity and likelihood each scored from 1 to 5. The resulting risk score is used to determine whether a risk is acceptable, requires further mitigation, or must be escalated.

Risks are evaluated both before and after the application of risk controls. The residual risk score is used to determine whether a risk is acceptable for release.

Risks with a residual score of 12 or above are escalated to the SCRUM MANAGER.

Clinical Risk Control
The table below shows how risks are rating prior to sufficient control measures.

Initial Risk Rating	Number of Risks
High	0
Medium	0
Low	0

Residual Risk Evaluation
Following confirmation of completion of control actions in the hazard log, the number of residual risks is recorded.

Residual Risk Rating	Number of Risks
High	0
Medium	0
Low	0

Risk Review
Clinical risks associated with Suvera Recall, all use with acceptable residuals are reviewed by the Clinical Safety Officer (CSO) at least annually, or sooner if triggered by incidents, software changes, or new hazards. Over-based, **AND SET** updates needed to risk controls, assessments, or the risk register are made and recorded.

Risk reviews are also triggered by changes to the system, clinical guidance, or relevant safety incidents. Any changes to risk controls or residual risk ratings are documented in the Clinical Risk Register, including the rationale behind decisions made. Where a risk is deemed unacceptable, it is escalated to the Clinical Safety Officer (CSO) and may require additional mitigation before the system can be released or continue in use.

Test Issues
Testing of risk control measures forms part of the overall testing strategy for Suvera Recall. Tests are currently, or explicitly test key risk controls, including those relating to clinical safety, data integrity, and system availability. Test results are reviewed before any release that includes changes to risk controls.

The testing covers all aspects of:

- Functionality testing (ensuring the system works as intended, e.g., without errors or unexpected behaviour)
- Performance testing (assessing how the system performs under various conditions, such as high user load or large data volumes)
- Security testing (identifying vulnerabilities and ensuring the system is protected against threats)
- Resilience testing (tests how the system recovers from failures, such as crashes or data loss)

Any issues found during testing are logged, tracked, and resolved before release. Regression testing is carried out to ensure that changes do not introduce new risks.

Medical Device Status
Suvera Recall is not a medical device and does not control, monitor, or process data for the Medicines and Healthcare products Regulatory Agency (MHRA). Suvera Recall is not a medical device under the Medical Devices Regulations 2002 (as amended). It does not control, monitor, or process data for the Medicines and Healthcare products Regulatory Agency (MHRA). It is classified as a non-medical software system.

Algorithms, logic-based coding that identifies action or recommended investigations, or treatments. It does not provide diagnostic, therapeutic, or monitoring functionality. Its role is limited to supporting workflow, reminders, and record-keeping. Clinical decision-making remains entirely with the clinician. Suvera Recall is not intended to be used as a medical device and is not marketed as such.

Safety Incidents
Safety incidents are defined as any event or near miss that could have led to harm to a patient, or that has actually resulted in harm, where the system may have contributed.

Incidents are reported internally by users and escalated to the Clinical Safety Officer (CSO) for investigation. Where appropriate, incidents are also reported externally in accordance with applicable regulatory requirements.

Each incident is investigated to determine its cause, contributing factors, and impact. Findings are used to update the clinical risk register, risk controls, and system documentation where necessary.

Conclusion
Based on the clinical risk management activities described in this report, the Clinical Safety Officer (CSO) considers that all identified clinical risks associated with Suvera Recall have been reduced to an acceptable level.

All high and medium residual risks have been reviewed and accepted, and appropriate risk controls are in place. The system is considered safe for release and ongoing use, subject to the monitoring and review activities described in this report.

Appendix A
Definitions
Definitions of terms used in the software based through the hazard log and across the CSCR. Definitions are consistent with ISO 14971 and DCB0129 where applicable. This is to be re-assessed.

Risk Status
Status	Definition
Active	Clinical risk remains unresolved or requires ongoing monitoring and mitigation
Monitored	Risk is controlled but is reviewed periodically, with additional controls not currently required
Transferred	Risk is managed by a third party, typically through contractual terms or deployment of the system. Further risk management will be undertaken by the third party.

Likelihood
Likelihood	Definition
Very High	Expected to occur almost certainly, highly likely to occur
High	Likely to occur frequently, probable, reasonably expected to occur at the time
Medium	Possible
Low	Unlikely to occur in the past majority of occasions will not
Very Low	Negligible or nearly negligible possibility of occurrence

Consequence
Consequence	Definition
Catastrophic	Death or permanent harm ranging from significant to total disability
Major	Permanent lifelong condition, or death of a person who has been receiving long-term treatment or care for a condition
Moderate	Injury or illness requiring professional intervention, that results in a substantial temporary reduction in quality of life lasting more than 7 days, but not permanent
Minor	Injury or illness requiring minor intervention, such as first aid treatment, that results in a temporary reduction in quality of life lasting less than 7 days
Negligible	Minimal injury or illness requiring no treatment, with no reduction in quality of life

Risk Matrix
Likelihood \ Consequence	Negligible	Minor	Moderate	Major	Catastrophic
Very High	5	10	15	20	25
High	4	8	12	16	20
Medium	3	6	9	12	15
Low	2	4	6	8	10
Very Low	1	2	3	4	5

Risk Acceptability Criteria
Risk Score	Acceptability
1–4	Acceptable
5–9	Tolerable with monitoring
10–14	Undesirable, mitigation required
15–25	Unacceptable, must be reduced